

DECLARATION OF CONFORMITY

MANUFACTURER: **KOKEN CO., LTD.**
1-4-14 Koraku, Bunkyo-ku, Tokyo 112-0004 Japan
Phone:+81-03-3816-3542 Fax:+81-03-3816-3582
URL:<http://www.kokenmpc.co.jp>
<mailto:customer@kokenmpc.co.jp>

EUROPEAN AUTHORISED REPRESENTATIVE: **MDSS GmbH**
Schiffgraben 41, D-30175 Hannover, Germany
Phone:+49-511-6262-8630 Fax:+49-511-6262-8633
www.mdssar.com

PRODUCT: Neonatal Resuscitation Simulator-Advance
LM-111GM

We, the manufacturer, herewith declare that the products mentioned above meet the provisions of the RE Directive 2014/53/EU and RoHS 2 Directive 2011/65/EU. All supporting Documentation is retained under the premises of the manufacturer.

STANDARDS APPLIED:

RE Directive:

Article 3.1a

EN 62368-1: 2014+A11: 2017
EN 62479: 2010

Article 3.1b

EN 301 489-1 V2.1.1(2017-02)
EN 301 489-17 V3.1.1(2017-02)
EN 61326-1:2013

Article 3.2

EN 300 328 V2.2.2(2019-07)

RoHS Directive:

EN IEC 63000 : 2018

START of CE-MARKING: **30/3/2021**

PLACE, DATE of ISSUE: **Yamagata, Japan, 30/3/2021**

Signature : 
Name Kenji Fukushima
Position Head of Tsuruoka Plant